


User Guide

ImegenTM Pollock ID kit

Theragra Chalcogramma detection
by real time PCR

REF : IMG-222



Instituto de Medicina Genómica SL
Agustín Escardino 9,
Parc Científic de la Universitat de València
46980 Paterna (Valencia, España)
+34 963 212 340 - info@imegen.es

imegen.es

Rev. 2

The information in this guide is subject to change without notice.

*Imegen guarantees that its products are free from defects, both in used materials as in its manufacturing process. This warranty is extended to the expiration date, as long as the storage conditions specified in this manual are met. **Our products are designed for use in research.** The user of the product is responsible for validating the usefulness of the protocol proposed by Imegen. These protocols are considered a guide only. Imegen does not offer any other warranty, express or implied, which extend beyond the proper functioning of the components of this set. Imegen sole obligation in respect of the preceding guarantees, will be to replace the product or return the purchase price thereof, as desired by the customer, as long as the existence of a defect in the materials test, or in the manufacture of its products. Imegen will not be responsible for any damage, direct or indirect, resulting in economic losses or damages resulting from the use of this product by the purchaser or user.*

*All products sold by the Imegen are subjected to rigorous quality control. The **imegenTM Pollock ID kit** has passed all internal validation tests, ensuring the reliability and reproducibility of each assay.*

For any questions about the applications of this product or its protocols, please contact our Technical Department:

Phone number: 963 212 340

e-Mail: info@imegen.es

****imegenTM** is a trademark of Instituto de Medicina Genómica, registered as imegen[®] in Spain*

Index

1. <i>Product information</i>	4
<i>Kit description</i>	4
<i>Content and storage of the kit</i>	4
<i>Equipment and material required</i>	5
2. <i>Real time PCR</i>	7
<i>PCR reactions preparation</i>	7
<i>PCR amplification program</i>	8
3. <i>Analysis of results</i>	10

1. Product information

Kit description

Identification of meat species presence in food samples is an essential step in order to verify the origin and traceability of the used raw materials, as well as a necessary quality control for handling and cleaning processes of production lines.

imegen-Pollock ID Kit allows determining the presence of DNA of *Theragra chalcogramma* in any food.

Theragra chalcogramma DNA detection is done by real time PCR using TaqMan[®]-MGB probes. One of them, labelled with FAM[™] dye, specifically detects one mitochondrial DNA sequence of *Theragra chalcogramma*. The second probe is labelled with VIC[®] and detects an Internal Positive Control, which is used to rule out inhibitors in the sample and check the correct functioning of the assay.

*The detection limit has been calculated upon standard samples consisting of mixtures of raw *Theragra chalcogramma* meat and other species. **imegen-Pollock ID Kit** can detect blends containing a percentage over 0.01% of *Theragra chalcogramma* meat. The limit of detection in processed samples varies depending on the composition and food processing.*

*To ensure the representativeness of the results, we recommend the use of a DNA extraction method that allows you to process a large amount of sample [10-20 g]. If you do not have a procedure with these features, we recommend the use of **Food Extraction Kit** [Part No: IMG-262].*

Content and storage of the kit

The kit contains needed reagents to perform the real time PCR reactions:

- Amplification primers for the PCR system.
- Two probes labelled with FAMTM and VIC[®] fluorophores, to detect the specific species and the Internal Positive Control [IPC].
- Pollock Positive Control [0,1% of *Theragra chalcogramma*].
- 2X General Master Mix.

The kit contains the necessary reagents to perform 48 reactions with each PCR system:

Reagents	Color	Amount	Storage
Pollock Master-Mix	Green pad	360 µl	-20°C
Pollock positive control	Yellow cap	60 µl	-20°C
General Master-Mix	White pad	600 µl	4°C

Table 1. Kit components and storage temperature of imegen-Pollock ID Kit

This kit has been proved and is compatible with the following real time PCR platforms: 7500 FAST and StepOne Real-Time PCR System [Thermo Scientific].

Equipment and material required

In the following table the equipment requirements for using imegen-Pollock ID Kit are shown:

Equipments	
1	Real-time PCR Thermal Cycler with channels for detection of FAM TM (520 nm) and VIC [®] (550 nm)
2	Micropipettes (10 µL, 20 µL and 200 µL)
3	Table top centrifuge with adaptors for 96 well PCR plates and/or 0,2 ml tubes
4	Vortex

Materials	
1	Optical 96-well reaction plates or 0.2 ml optical tubes
2	Optical adhesive film for 96 well plates or optical caps for 0.2 ml tubes
3	Disposable micropipette filter tips (10 μ L, 20 μ L and 200 μ L)
4	1.5 mL sterile tubes
5	Powder-free latex gloves

2. Real Time PCR

PCR reactions preparation

A PCR master mix should be prepared in order to perform the analysis with *imegen-Pollock ID kit*:

The PCR Master Mix contains:

- Pollock Master-Mix
- General Master-Mix (2X)

The recommended protocol for preparation of amplification reactions is showed below:

1. Thaw the Pollock Master-Mix, Pollock Positive Control and samples DNA.
2. Vortex each reagent and keep cold.
3. Add into a 1.5 mL tube, one for each PCR master mix preparation, the following reagents [table 2]. To estimate the amount of necessary reagents, we recommend make calculations taking into account the number of samples to be simultaneously analysed, and then considering one more reaction, or increase a 10% the volume of each reagent.

Reagents	Amount per reaction
Pollock Master-Mix	7.5 µL
General Master Mix	12.5 µL

Table 2. Reagents amount per reaction in 7500 FAST or StepOne

4. Vortex the PCR Master Mixes tubes and dispense 20 µL into corresponding wells.
5. Add 5 µl of each DNA sample at 10 ng/µl, 5 µl of Pollock Positive Control and 5 µl of the Negative Controls* into the appropriate wells.
We recommend making each sample analysis in duplicate.
6. Cover the well plate with optical film or the tubes with optical cover and spin in the centrifuge.

* We strongly recommend using an **extraction negative control** for each run of extractions carried out. This control consists in one tube to which no sample is added and which is submitted to the same extraction process as the other samples. Likewise, we recommended using a **PCR negative control** for each PCR run; this tube contains no DNA but all PCR reagents.

PCR amplification program

This kit is compatible with the Real-time PCR platforms 7500 FAST, StepOne Real-Time PCR System [Thermo Scientific], and any other Real-time PCR platforms equipped with FAM and VIC channels.

Probes	Receptor	Quencher
Pollock	FAM TM	MGB
IPC-P	VIC [®]	MGB

Table 4. Probes information

The following instructions should be taken into account in order to setup the amplification program:

- Reaction volume: 25 µL
- Targets: FAM and VIC

- In case the quencher has to be defined, select MGB for all probes. If the real time PCR system does not take into account the quenchers, select only the receptors [FAM and VIC].
- If the real time PCR system is a 7500 Fast or a StepOne Real-Time PCR system [Thermo Scientific] select *Quantitation- Standard curve* as a type of experiment and include ROX™ as a reference.
- Ramp rate: *standard*
- Optimal program:

Fields	Step 1 Enzyme activation	Step 2 PCR	
Cycle Number	1 initial cycle	40 cycles	
		Denaturation	Primers binding/Extention
Temperature	95°C	95°C	60°C
Time	10 minutes	15 seconds	1 minute*

Table 5. Optimal PCR program

*Fluorescence detection

Note: This program has been validated on a StepOne Real-Time PCR System from Applied Biosystems. If you use another brand or model of thermal cycler, you may need the amplification program to be adjusted. Please contact our service department for advice.

3. Analysis of results

Before analysing the samples results, you should establish the threshold in 0.4 and, also, you should check that the results obtained in the controls, is as expected:

- **Positive Control:** The result must always be positive in all amplification reactions, both in the FAMTM channel as VIC[®].
- **Negative controls:** Amplification must be only detected in the VIC[®] channel. In this channel an internal positive control [IPC] is detected, which determines the absence of inhibition in the sample.

IPC

It must be checked that the IPC [VIC[®]] is positive in all samples, with a Ct similar to the Positive Control. A negative result in the IPC indicates the presence of inhibitors in the sample. It should be noted that IPC result may be negative in samples where a lot of *Theragra chalcogramma* DNA [FAMTM] is detected, because the PCR reagents are exhausted before amplification of the IPC begins.

Pollock

Amplification in the FAMTM channel indicates presence of *Theragra chalcogramma* DNA in the sample.

It is necessary to check if sample Ct is less than the Ct_{cut-off} in order to determine if one reaction of amplification is positive. Any reaction of amplification with Ct upper than Ct_{cut-off} may be considered as negative. The Ct_{cut-off} is equal than the positive control Ct [0.1%] plus 3.32.

$$Ct_{cut-off} = 3.32 + Ct_{Positive Control}$$

*Note: Any sample with a Ct equal than Ct_{cut-off} contains approximately 0.01% of *Theragra chalcogramma* DNA.*

In samples where no amplification in the FAMTM channel is seen, we can conclude that no *Theragra chalcogramma* DNA is detected or that their amount in the sample is below than the detection limit.

The following table shows graphically the results that may be obtained from one sample analysis, as well as the interpretation that should be done from the obtained result:

Pollock Master Mix		Interpretation
Pollock	IPC	
-	+	No <i>Theragra chalcogramma</i> DNA are detected
+	+	<i>Theragra chalcogramma</i> is detected
-	-	PCR inhibitors presence in the sample*
+	-	Sample with big amount of <i>Theragra chalcogramma</i> DNA

Table 5. Results interpretation

* If presence of inhibitors in the sample is detected, we recommend checking whether there has been an excess of DNA in the reaction [the recommended maximum is 250 ng]. If the amount of DNA is right, we recommend repeating DNA extraction. If the problem persists, please contact our technical department.

The following table shows graphically the results that may be obtained from the analysis of different assay controls, as well as the interpretation that should be done from the obtained result:

Controls	Pollock Master Mix		Interpretation
	Pollock	IPC	
Positive control	+	+	Expected result
	-	-	PCR Amplification Failure ¹
Extraction negative control	-	+	Expected result
	+	+	Contamination in the DNA extraction procedure ²
PCR negative control	-	+	Expected result
	+	+	PCR contamination with Theragra chalcogramma DNA ³

Table 6. Interpretation of possible results

■ **Recommendations:**

¹ **PCR Amplification Failure:** Check amplification program and configuration of fluorescence capture. Amplification failure may be due to a setup technical problem.

² **Contamination in the DNA extraction procedure:** Contamination may be due to some error made in the process of sample handling, reagents contamination, or environmental contamination. Check DNA extraction protocol, wipe the laboratory where DNA extraction process was performed and take care to avoid any contamination during sample homogenization. If necessary, use new aliquots of the reagents used in DNA extraction.

³ **PCR contaminations with Theragra chalcogramma DNA:** Contamination of PCR reactions may be due to an error made in the process of sample handling, contamination of the reagents or environmental contamination. Thoroughly clean the laboratory where the PCR process was performed, as well as equipment. If necessary, use new aliquots of the reagents used in the PCR. Prepare the PCR reaction containing the Positive Control last to avoid cross contamination.