



Instructions for Use

Imegen[®] Celery Allergen ID kit

REF

IMG-180

Manufactured by:

Health in Code, S.L

Calle de la Travesía s/n, 15E Base 5, Valencia 46024, España

+34 963 212 340 - info@healthincode.com

imegen.es



The information in this guide is subject to change without notice.

Health in Code S.L. guarantees that its products are free from defects, both in used materials as in its manufacturing process. This warranty is extended to the expiration date, as long as the storage conditions specified in this manual are met. **Our products are designed for research use only.** The user of the product is responsible for validating the usefulness of the protocol proposed by Health in Code S.L. These protocols are considered a guide only. Health in Code S.L. does not offer any other warranty, express or implied, which extend beyond the proper functioning of the components of this set. Health in Code S.L. sole obligation in respect of the preceding guarantees, will be to replace the product or return the purchase price thereof, as desired by the customer, as long as the existence of a defect in the materials test, or in the manufacture of its products. Health in Code S.L. will not be responsible for any damage, direct or indirect, resulting in economic losses or damages resulting from the use of this product by the purchaser or user.

All products sold by the Health in Code S.L. are subjected to rigorous quality control. The **Imegen[®] Celery Allergen ID kit** has passed all internal validation tests, ensuring the reliability and reproducibility of each assay.

For any questions about the applications of this product or its protocols, please contact our Technical Department:

Phone number: +34 963 212 340

e-mail: tech.support@healthincode.com

Imegenagro[®] is a trademark of Health in Code S.L.

Amendments to the Instructions for Use (IFU)	
Version 05	Amendment: General Master Mix temperature update.
Version 04	Amendment: Change of kit's name, "Allergen" in capital letter and manufacturer's name and address update.
Version 03	Amendment: Change of the manufacturer's identification, going from Imegen to Health in Code.
Version 02	Amendment: Contents review



Index

1	General Information	4
2	Intended Use	5
3	Technical Characteristics	5
4	Warnings and precautionary statements	7
5	Contents and Storage Conditions	7
6	Equipment and Material required but not supplied	8
7	Assay Protocol	9
	7.1 PCR reactions preparation	9
	7.2 Configuration of the PCR program in real time	10
8	Results Analysis	12
9	Troubleshooting	14
10	Limitations	15
	10.1 Equipment	15
	10.2 Reagents	15
	10.3 Product Stability	15

1. General information

A considerable proportion of the general public experiences negative health effects triggered by certain components contained in their habitual diet. Susceptible persons can develop an intolerance (e.g. against lactose, due to the lack of a digestive enzyme), and others suffer from adverse reactions mediated by the immune system. The latter is characterised by the action of IgE antibodies against the offending food and is known as food allergy. The symptoms caused by allergic reactions range from rather mild manifestations to life threatening events (anaphylactic shock).

Usually, the recommended option to treat food allergies is to eliminate from the diet the food ingredients that cause the hypersensitivity. Allergic persons need to know whether the food items they purchase contain allergenic ingredients; they have to rely on the truthfulness of information given on the label of prepared and packaged food items. National and supra-national legislation has been put in place which requires food business operators to declare whether ingredients with a known allergenic potential have been used during manufacturing. The Codex Alimentarius General Standard for the Labelling of Prepacked Food requires, for example, the mandatory labelling of the presence of eight classes of food ingredients that are known to cause hypersensitivity in susceptible consumers (cereals containing gluten, crustaceans, eggs, fish, peanuts, soybeans, milk, and tree nuts), while European Community legislation (Commission Directive 2007/68/EC) extends the list to include also celery, mustard, sesame seeds, lupin, and molluscs.

Analytical testing systems are needed by the food industry to enable them to test whether allergens are present in their raw materials, the finished products and whether production lines have been correctly sanitised, by the food inspection authorities for market surveillance and by academia to enable and stimulate research into food allergy and allergen detection. Molecular biology provides highly specific and sensitive procedures to detect the presence of allergen species.



In order to preserve the health of consumers, Health in Code S.L. has been developed some kits to detect plant species and food allergens based on the latest DNA technology, used to verify the presence or absence of allergenic ingredients in any food.

References:

Paschke, A. & Ulberth, F. Anal Bioanal Chem (2009) 395: 15. <https://doi.org/10.1007/s00216-009-2989-0>

2. Intended Use

In order to preserve the health of consumers, it has been developed some kits to detect plant species and food allergens based on the latest DNA technology, used to verify the presence or absence of allergenic ingredients in any food.

Nowadays, celery is widely cultivated in temperate regions around the world, particularly in Europe and North America. And its use is becoming more common and common.

Identification of celery presence in food samples is an essential step in order to verify the origin and traceability of the used food, as well as a necessary quality control for handling and cleaning processes of production lines.

3. Technical characteristics

Imegen® Celery Allergen ID Kit allows determining the presence of DNA of celery (*Apium graveolens*) in any food.

Celery DNA detection is done by real time PCR using two hydrolysis probes. One of them, labelled with FAM™ dye, specifically detects one chloroplast DNA sequence of celery. The second probe is labelled with VIC™ and detects an Internal Positive Control (IPC), which is used to rule out inhibitors in the sample and check the correct functioning of the assay.

To ensure the representativeness of the results, we recommend the use of a DNA extraction method that allows you to process a large amount of sample (10-20 g). If you do not have a procedure with these features, we recommend the use of **Food Extraction Kit** (Part No: IMG-262).

The limit of detection of the assay is 0.0001% (equivalent to 1 ppm).

The specificity of the kit was tested through comparison with the NCBI sequence database and



was also experimentally tested with success on a collection of reference DNAs.

Health in Code S.L. is certified by BV against the standard UNE-EN ISO 9001 "Quality management systems" for the design, development, manufacture and commercialization of kits for genetic analysis (Certification number ES090493-1)

4. Warnings, precautionary statements and requirements

1. Strictly follow the instructions of this manual, especially regarding the handling and storage conditions.
2. Do not pipette by mouth.
3. Do not smoke, drink, or eat in areas where specimens or kit reagents are being handled.
4. You must properly protect any skin condition, as well as cuts, abrasions and other skin lesions.
5. Send down the drain only those materials found on the safe list. Compounds not listed are not suitable for drain disposal. Use waste containers according to the local legislation and manage their treatment through an authorised waste manager.
6. In case of an accidental release of any of the reagents, avoid contact with skin, eyes and mucous membranes and clean with abundant water.
7. The materials safety data sheets of all hazardous components contained in this kit are available on request to Health in Code S.L.
8. This product could require the handling of samples and materials of human and animal origin. You should consider all human and animal source materials as potentially infectious and handled in accordance with OSHA Biosafety Level 2 of bloodborne pathogens or must use other relevant biosafety practices for materials containing or suspect that they may contain infectious agents.
9. Reagents included in this kit are non-toxic, neither explosive, infectious, radioactive, magnetic, corrosive nor environmental pollutants.
10. This kit has been validated with specific equipment under certain conditions, which could be different in other laboratories. It is recommended that each laboratory performs an internal validation when the kit is used for the first time.
11. The manufacturer is not responsible for the malfunction of the assay when one or more reagents included in the kit are replaced by other reagents not supplied by Health in Code S.L.
12. The manufacturer does not guarantee the reproducibility of the assay when the user employs reagents not validated by Health in Code S.L., considering them equivalent to those provided in the Kit.

5. Content and storage conditions

The kit contains the necessary reagents to perform 48 reactions.

- **Celery Master Mix:** Master Mix with specific oligonucleotides, fluorophore-labelled hydrolysis probes (FAM™ probe for celery detection and VIC™ probe for the Internal Positive Control detection, IPC), synthetic plasmid including the specific IPC sequence and nuclease-free water.
- **General Master Mix:** Master Mix of PCR with nucleotides, MgCl₂, DNA polymerase and buffer needed to carry out real-time PCR.
- **Positive Control:** DNA containing a 0.1% of *Apium graveolens* specific-DNA

Reagents	Colour	Amount	Storage
Celery Master Mix	Green pad	360 µL	-20°C
General Master Mix	White pad	600 µL	-20°C upon receipt. 2–8°C after first use. Keep protected from light.
Positive control	Green cap	60 µL	-20°C

Table 2. Kit components and storage temperature of Imegen® Celery Allergen ID Kit

6. Equipment and material required but not supplied

Equipment:

- Real-time PCR Thermal Cycler with channels for detection of FAM™ (520 nm) and VIC™ (550 nm)
- Micropipettes (10 µL, 20 µL and 200 µL)
- Table top centrifuge with adaptors for 96 well PCR plates and/or 0,2 ml tubes
- Vortex

Materials:

- Optical 96-well reaction plates or 0.2 ml optical tubes
- Optical adhesive film for 96 well plates or optical caps for 0.2 ml tubes
- Disposable micropipette filter tips (10 µL, 20 µL and 200 µL)
- 1.5mL sterile tubes
- Powder-free latex gloves

Reagents:

- Nuclease-free water

7. Assay Protocol

7.1 PCR reactions preparation

Imegen® Celery Allergen ID Kit is designed to determine, in a single PCR reaction, the presence or absence of celery DNA and the internal positive control.

We recommend using, the positive control included in this kit for each run.

The recommended protocol for preparation of amplification reactions is showed below:

1. Thaw the Celery Master Mix and the Positive Control vial.
2. Vortex each reagent and keep cold.
3. Add into a 1.5 mL tube (one of each PCR master mix preparation), the following reagents (Table 3). To estimate the amount of necessary reagents, we recommend make calculations taking into account the number of samples and controls to be simultaneously analyzed, and then considering one more reaction, or increase a 10% the volume of each reagent.

Reagents	Amount per reaction
Celery Master Mix	7.5 µL
General Master Mix	12.5 µL

Table 3. Reagents amount per reaction

4. Vortex and spin the 1.5 mL tube and dispense 20 µl per well or tube of 0.2 ml.
5. Add 5 µl of each DNA sample at 10 ng/µl, 5 µl of Pollock Positive Control and 5 µl of the Negative Controls* into the appropriate wells.
6. Cover the well plate with optical film or the tubes with optical cover and spin in the centrifuge.

** We strongly recommend using an **extraction negative control** for each run of extractions carried out. This control consists in one tube to which no sample is added and which is submitted to the same extraction process as the other samples. Likewise, we recommended using a **PCR negative control** for each PCR run; this tube contains no DNA but all PCR reagents.*



7.2 PCR amplification programme

This kit is compatible with the Real-time PCR platforms *7500 FAST*, *StepOne Real-Time PCR System* (Thermo Scientific), and any other Real-time PCR platforms equipped with FAM™ and VIC™ channels.

Probes	Receptor	Quencher
Celery-P	FAM TM	MGB
IPC-P	VIC TM	MGB

Table 4. Probes information

The following instructions should be taken into account in order to setup the amplification program:

- Reaction volume: 25 µL
- Targets: FAMTM and VICTM
- In case the quencher has to be defined, select MGB for all probes. If the real time PCR system does not take into account the quenchers, select only the receptors (FAMTM and VICTM).
- If the real time PCR system is a 7500 Fast or a StepOne Real-Time PCR system (Thermo Scientific) select *Quantitation- Standard curve as a type of experiment and include ROXTM* as a reference.
- Ramp rate: *standard*

- Optimal program:

Fields	Step 1	Step 2	
	Enzyme activation	PCR	
Cycle Number	1 initial cycle	50 cycles	
		Denaturation	Primers binding/Extention
Temperature	95°C	95°C	60°C
Time	10 minute	15 seconds	1 minute*

Table 5. Optimal PCR program *Fluorescence detection

■ **Analysis:**

To analyse the results, it is recommended to establish the Threshold at 0.1 and to keep the default Baseline value to minimise the residual signal in the detection channels.

Ct Settings:

Threshold	0.1
Baseline	AUTO

8. Results analysis

Before analysing the samples results, it should be checked if obtained results in controls are as expected:

- **Positive Control:** The result must always be positive in all amplification reactions, both in the FAMTM channel as VICTM.
- **Negative controls:** Amplification must be only detected in the VICTM channel. In this channel an internal positive control (IPC) is detected, which determines the absence of inhibition in the sample.

IPC

It must be checked that the IPC (VICTM) is positive in all samples, with a Ct similar to the Positive Control. A negative result in the IPC indicates the presence of inhibitors in the sample. It should be noted that IPC result may be negative in samples where a lot of celery DNA (FAMTM) is detected, because the PCR reagents are exhausted before amplification of the IPC begins.

Celery

Amplification in the FAMTM channel indicates presence of celery DNA in the sample.

In samples where no amplification in the FAMTM channel is seen, we can conclude that no celery DNA is detected or that its amount in the sample is below than the detection limit.

The following table shows graphically the results that may be obtained from one sample analysis, as well as the interpretation that should be done from the obtained result:

Celery Master Mix		Interpretation
Celery	IPC	
-	+	No celery DNA is detected
+	+	Celery DNA is detected
-	-	PCR inhibitors presence in the sample*
+	-	Sample with big amount of celery DNA

Table 6. Results interpretation

** If presence of inhibitors in the sample is detected, we recommend checking whether there has been an excess of DNA in the reaction (the recommended maximum is 250 ng). If the amount of DNA is right, we recommend repeating DNA extraction. If the problem persists, please contact our technical department.*

9. Troubleshooting

The following table shows graphically the results that may be obtained from the analysis of different assay controls, as well as the interpretation that should be done from the obtained result:

Controls	Celery Master Mix		Interpretation
	Celery	IPC	
Positive control	+	+	Expected result
	-	-	PCR Amplification Failure ¹
Extraction Negative Control	-	+	Expected result
	+	+	Contamination in the celery DNA extraction procedure ²
PCR Negative Control	-	+	Expected result
	+	+	PCR contamination with celery DNA ³

Table 7. Possible results and their interpretation

¹ **PCR Amplification Failure:** Check amplification program and configuration of fluorescence capture. Amplification failure may be due to a setup technical problem.

² **Contamination in the celery DNA extraction procedure:** Contamination may be due to some error made in the process of sample handling, reagents contamination, or environmental contamination. Check DNA extraction protocol, wipe the laboratory where DNA extraction process was performed and take care to avoid any contamination during sample homogenization. If necessary, use new aliquots of the reagents used in DNA extraction.

³ **PCR contaminations with celery DNA:** Contamination of PCR reactions may be due to an error made in the process of sample handling, contamination of the reagents or environmental contamination. Thoroughly clean the laboratory where the PCR process was performed, as well as equipment. If necessary, use new aliquots of the reagents used in the PCR. Prepare the PCR reaction containing the Positive Control last to avoid cross contamination.

10. Limitations

10.1 Equipment

Imegen® Celery Allergen ID kit has been validated using the following real-time PCR systems:

- StepOnePlus™ Real-Time PCR System (ThermoFisher Scientific)
- 7500 FAST Real-Time PCR System (ThermoFisher Scientific)
- StepOne™ Plus Real-Time PCR System (ThermoFisher Scientific)

Technically, this kit is compatible with any real-time PCR systems that enable the detection of the fluorescence emitted by FAM™ and VIC™ fluorophores.

If a PCR system different from the systems described in this section is going to be used, it is possible that the PCR program might need to be readjusted. In this case, please contact our Technical Support Team for more details.

10.2 Reagents

Imegen® Celery Allergen ID kit has been validated using the reagents included in the kit and the DNA polymerase recommended by the supplier of the real-time PCR systems used in the validation as follows:

- TaqMan Environmental Master Mix 2.0 (ThermoFisher Scientific)

If a PCR master mix (DNA polymerase) different from the DNA polymerase used in the validation is going to be used to perform the analysis, a validation with the new reagents is recommended beforehand. Please, contact our Technical Support Team if you request any further information.

10.3 Product Stability

The optimal analytical functioning of this product is confirmed as long as the recommended storage conditions are applied as specified on Section 5 (Contents and Storage Conditions) from the reception of the kit until the expiry date assigned to each production batch.